


Mackworth CPP Portable Lift Troubleshooting Guide

Scenario: Battery Voltage Delta Exceeded

When operating the hand controls, the portable lift LED indicates a yellow and blue light and an audible alarm will repeat 4 beeps (0.5s apart) x2 cycles.



When this scenario occurs, it means that either one or both batteries aren't holding charge as expected

Replace the batteries as a pair

Key:

Question

Action

Component
Replacement

Supporting
Document

Additional Context
and Actions